

Table S1. Analytical parameters for method validation in multi-element analysis.

| Pine Needles Nist 1575 | | | | | | |
|------------------------|-----------------------------|-----------------------------|--------|-------------------------------|-------------------------------------|-----------------|
| Element | LOD ($\mu\text{g/Kg}$) | LOQ ($\mu\text{g/Kg}$) | R^2 | Experimental Value (mg/kg) | Linear Range ($\mu\text{g/L}$) | Recovery (%) |
| Li | 0.001 | 0.003 | 0.9998 | 1.95 * | 0.5–50.0 | 97.50 |
| B | 0.003 | 0.010 | 0.9997 | 9.1 | 0.5–50.0 | 94.79 |
| Na | 0.461 | 1.521 | 0.9982 | 60.00 | 0.5–50.0 | 95.24 |
| Mg | 0.031 | 0.102 | 0.9992 | 0.103 | 0.5–50.0 | 97.17 |
| Al | 0.021 | 0.069 | 0.9996 | 566 | 0.5–50.0 | 97.59 |
| K | 0.142 | 0.469 | 0.9989 | 0.380 | 0.5–50.0 | 91.13 |
| Cr | 0.001 | 0.003 | 0.9998 | 0.38 | 0.5–50.0 | 95.00 |
| Mn | 0.002 | 0.007 | 0.9997 | 478 | 0.5–50.0 | 97.95 |
| Fe | 0.021 | 0.069 | 0.9996 | 43.5 | 0.5–50.0 | 94.57 |
| Ni | 0.001 | 0.003 | 0.9998 | 1.43 | 0.5–50.0 | 97.28 |
| Cu | 0.014 | 0.046 | 0.9993 | 2.7 | 0.5–50.0 | 96.43 |
| Zn | 0.027 | 0.089 | 0.9992 | 37.45 | 0.5–50.0 | 98.55 |
| As | 0.001 | 0.003 | 0.9997 | 0.040 | 0.5–50.0 | 102.56 |
| Se | 0.042 | 0.139 | 0.9991 | 0.095 | 0.5–50.0 | 95.96 |
| Sr | 0.003 | 0.010 | 0.9997 | 1.99 * | 0.5–50.0 | 99.50 |
| Ca | 0.002 | 0.007 | 0.9998 | 2.00 * | 0.5–50.0 | 100.00 |
| Cd | 0.001 | 0.003 | 0.9998 | 0.225 | 0.5–50.0 | 96.57 |
| Sb | 0.002 | 0.007 | 0.9997 | 1.95 * | 0.5–50.0 | 97.50 |
| Ba | 0.004 | 0.013 | 0.9997 | 6.08 | 0.5–50.0 | 101.33 |
| Pb | 0.002 | 0.007 | 0.9998 | 0.170 | 0.5–50.0 | 101.80 |
| Hg | 0.001 | 0.003 | 0.9998 | 0.0404 | 1.0–100 | 101.25 |

* Not present in the certified matrix, added later to the matrix.