

Analytical method validation

The instrumental precision, sample repeatability and within-day stability were determined prior to the analysis of experimental samples. Quality control (QC) samples were obtained by mixing an equal amount (100 μ l of brain homogenates or 10 μ l of plasma) from each sample. The extracted ion chromatographic peaks of six ions (m/z 132.0762_0.94 min, m/z 137.0447_1.28 min, m/z 205.0979_6.70 min, m/z 161.0946_9.30 min, m/z 327.0093_12.53 min, m/z 526.2928_13.67 min) from brain samples and six ions (m/z 112.0498_1.13 min, m/z 205.0961_6.65 min, m/z 531.7684_7.73 min, m/z 318.2992_11.41 min, m/z 496.3388_14.52 min, m/z 524.3700_16.77 min) from plasma samples were selected for method validation. The selected ions were distributed over the analysis m/z and time range. The instrumental precision was validated by analyzing five injections continuously of the same QC sample. Then five aliquots of a random sample were used to investigate the sample repeatability. The within-day stability was evaluated using five injections of the same QC sample at 0, 4, 8, 12 and 24 h. The results (RSD%) of the retention times were all not more than 5% and the intensities were not more than 10%. The method validation results confirmed the reliability of the method.