

# Display Report

## Analysis Info

Analysis Name D:\Data\Chizhov\Krayushkin\Melekhina\01\_&clb.d  
Method tune\_wide.m  
Sample Name /MVGR 01  
Comment CH3CN 100%, dil. 200, calibrant added

Acquisition Date 15.06.2018 14:47:31

Operator BDAL@DE  
Instrument / Ser# micrOTOF 10248

## Acquisition Parameter

Source Type	ESI	Ion Polarity	Positive	Set Nebulizer	0.4 Bar
Focus	Not active			Set Dry Heater	180 °C
Scan Begin	50 m/z	Set Capillary	4500 V	Set Dry Gas	4.0 l/min
Scan End	3000 m/z	Set End Plate Offset	-500 V	Set Divert Valve	Waste

