S1 of S1

Supplementary Materials: High Sensitivity Determination of TNF-a for Early Diagnosis of Neonatal Infections with a Novel and Reusable Electrochemical Sensor

Liangliang Li, Miaomiao Li, Wenwen Wang, Qian Zhang, Dongyun Liu, Xianghong Li and Hong Jiang

-	<u> </u>		
Oligonucleotides	Sequences		
Capture probe (T1)	5'-SH-TTGTCGCCGACTGCGCCATCC-3'		
Aptamer (T2)	5'-TTTTTTTTTTTTTTTGGTGGATGGCGCAGTCGGCGACAA-3'		
Sgnal probe (T3)	5'- SH-GGATGGCGCAGT-3'		

Table S1. Sequences of the designed oligonucleotides.

Table S2. Comparison of recently reported biosensors.

Method	Analyte	Line arrange	Detection limit	Reference
Fluorescence aptasensor	TNF-a	5–100 pg/mL		[1]
Electrochemical aptasensor	TNF-a	1–10 ³ pg/mL	1 pg/mL	[2]
Electrochemical sensing	TNF-a	0–60 pg/mL	2.07 pg/mL	[3]
Electrochemical aptasensor	TNF-a	1–1 × 104 pg/mL	0.7 pg/mL	This work

- Kongsuphol, P.; Ng, H.H.; Pursey, J.P.; Arya, S.K.; Wong, C.C.; Stulz, E.; Park, M.K. EIS-based biosensor for ultra-sensitive detection of TNF-*α* from non-diluted human serum. *Biosens. Bioelectron.* 2014, *61*, 274– 279.
- 2. Castro-López, V.; Elizalde, J.; Pacek, M.; Hijona, E.; Bujanda, L. A simple and portable device for the quantification of TNF-*α* in human plasma by means of on-chip magnetic bead-based proximity ligation assay. *Biosens. Bioelectron.* **2014**, *54*, 499–505.
- 3. Mazloum-Ardakani, M.; Hosseinzadeh, L.; Taleat, Z. Synthesis and electrocatalytic effect of Ag@Pt core-shell nanoparticles supported on reduced graphene oxide for sensitive and simple label-free electrochemical aptasensor. *Biosens. Bioelectron.* **2015**, *74*, 30–36.



© 2017 by the authors; licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons by Attribution (CC BY) license (http://creativecommons.org/licenses/by/4.0/).