

# BTS - 600 Prüfprogramm: fstCheck

Datum/Zeit : 20.02.2018 10:30:34  
Bediener : Digatron

Version : V 1.600.395

Programm : fstCheck

Titel : first CheckUpTC

Schritt	Label	Prozedur	Sollwert	Limit	Aktion	Registrierung
1		SET				Multi 0.1 V 0.1 A 1 min
2		SET		> 3.0 V < 1.0 V	UNT UNT	
3		SET	ClimaOn = 1.0			
4		SET	SetTemp = 25.0			
5		UNT				
6		PAU		1 min		
! first EIS measurement						
7		PAU		4:20 h		
8		SET	EISstart = 1.0			
9		PAU		0:40 h		
10		SET	EISstart = 0.0			
11		PAU		EISready > 0.0 5 min		
! first charging and cap-test						
12		LAD	5.0 ACN20 2.6 V	30 h		
13		BEG				
14		PAU		4 h		
15		ELA	4.0 ACN20	< 1.8 V		
16		PAU		4 h		
17		LAD	5.0 ACN20 2.6 V	24 h		
18		ZYK	3*			
! EIS at 80 % SoC						
19		PAU		24 h		
20		ELA	2.0 ACN20	2.0 AhStep		
21		EISwDCA				
! EIS at 40 % SoC						
22		ELA	2.0 ACN20	4.0 AhStep		
23		EISwDCA				
! EIS at 0 % SoC						
24		ELA	2.0 ACN20	4.0 AhStep		
25		EISwDCA				
! recharge						
26		LAD	5.0 ACN20 2.6 V	24 h		
27		PAU		10 h		
28		STO				