

## Validation report

### 1. Validation Method

#### 1.1. Linearity

Six calibration curves were prepared with eight calibration standards (CS). Quantification was achieved by plotting the peak area ratios of uremic toxins (UT) to their internal standards (IS), and a weighting factor was determined. Back-calculated concentrations of the CS have to be within 85–115% of the nominal concentrations, except for the lower limit of quantification (LLOQ) a deviation of 20% was considered acceptable.

#### 1.2. Lower Limit of Quantification (LLOQ)

The LLOQ was the lowest concentration achievable with an accuracy between 80 and 120 % and a coefficient of variation of  $\pm 20$  % obtained over six measurements at three different days. Three levels were evaluated: 25, 100 and 500 ng/mL.

#### 1.3. Carry over

A blank sample was analyzed immediately after running three replicates of the highest CS to determine the carryover.

#### 1.4. Accuracy and precision

The accuracy (bias, %) and precision (coefficient of variation, CV %) of the assay were determined for the three QC levels (25, 100 and 500 ng/mL) over three different days. Each day, 6 replicates of each QC were processed. The intra-assay bias and CV takes into account the variability of the 6 replicates within each validation day and the inter-assay bias and CV evaluates the variability between the three validation days. An accuracy within the range 85-115% of the nominal values and a precision with a CV of  $\pm 15\%$  were required.

### 2. Validation results

#### 2.1. Linearity

Table S1. Mean concentration of calibrant, bias, standard deviation and coefficient of variation.

		Theoretical concentrations (ng/mL)						
Molecules		1000	500	250	100	50	25	10
Nicotine	Mean (n=6)	1050	482	246	99	46	24	11
	Bias	-4,9%	3,7%	1,8%	1,3%	8,0%	5,9%	-5,4%
	SD	32,6	26,8	10,6	5,3	1,2	1,1	0,7
	CV	3,1%	5,6%	4,3%	5,4%	2,5%	4,7%	7,1%
Cotinine	Mean (n=6)	1033	483	232	101	48	24	10
	Bias	-3,3%	3,4%	7,0%	-1,3%	3,4%	4,8%	-2,8%
	SD	20,1	19,5	1,8	5,2	2,0	1,4	0,5
	CV	2,0%	4,0%	0,8%	5,2%	4,0%	5,7%	4,4%
3-hydroxycotinine	Mean (n=6)	1035	486	230	99	47	24	11
	Bias	-3,5%	2,8%	7,9%	0,9%	5,8%	5,2%	-8,0%
	SD	23,9	28,0	4,8	5,5	3,4	1,3	0,3
	CV	2,3%	5,8%	2,1%	5,5%	7,2%	5,5%	3,1%

### 2.2. Lower Limit of Quantification

Table S2. Lower limit of quantification (LLOQ).

Intra-day (n=6)			
Molecules	LLOQ (ng/mL)	CV (%)	Bias (%)
Nicotine	10	2.0%	-7.8%
Cotinine	10	2.2%	-3.7%
3-hydroxycotinine	10	3.6%	2.5%

### 2.3. Carry over

No chromatographic peaks were evident for the analytes and for the internal standards when a blank was injected just after the highest point in the calibration.

# 2.4. Accuracy and precision

Table S3. Intra-day and inter-day accuracy and precision.

Molecules	Theoretical concentration (ng/mL)	Intra-day repeatability (n = 6)		Inter-day repeatability (n = 6)	
		CV (%)	Bias (%)	CV (%)	Bias (%)
Nicotine	25	14%	-7.7%	12%	-2.0%
	100	13%	-7.6%	9.9%	1.8%
	500	10%	1.1%	8.0%	2.4%
Cotinine	25	13%	7.8%	6.8%	0.6%
	100	4.0%	9.7%	9.5%	1.3%
	500	7.4%	5.9%	5.5%	2.4%
3-hydroxycotinine	25	11%	-2.8%	5.6%	-7.2%
	100	3.6%	-12%	7.2%	-1.0%
	500	8.1%	-0.3%	13%	0.9%