

Table S1. Results of first and repeated 10 IGF-1 samples measured by radioimmunoassay stored >9 years at -80°C.

Month/year of first measurement	Value (ng/ml)	Month/year of second measurement	Value (ng/ml)	Mean (ng/ml)	SD	CV (%)
<u>11/2011</u>	<u>1813</u>	<u>02/2021</u>	<u>1533</u>	<u>1673</u>	<u>140</u>	<u>8.37</u>
<u>06/2011</u>	<u>1736</u>	<u>02/2021</u>	<u>1779</u>	<u>1757.5</u>	<u>21.5</u>	<u>1.22</u>
<u>06/2011</u>	<u>1321</u>	<u>02/2021</u>	<u>1356</u>	<u>1338.5</u>	<u>17.5</u>	<u>1.31</u>
<u>12/2011</u>	<u>1302</u>	<u>02/2021</u>	<u>1591</u>	<u>1446.5</u>	<u>144.5</u>	<u>9.99</u>
<u>01/2010</u>	<u>1907</u>	<u>02/2021</u>	<u>2001</u>	<u>1954</u>	<u>47</u>	<u>2.41</u>
<u>10/2011</u>	<u>1531</u>	<u>02/2021</u>	<u>1579</u>	<u>1555</u>	<u>24</u>	<u>1.54</u>
<u>01/2012</u>	<u>1442</u>	<u>02/2021</u>	<u>1761</u>	<u>1601.5</u>	<u>159.5</u>	<u>9.96</u>
<u>01/2010</u>	<u>842</u>	<u>02/2021</u>	<u>857</u>	<u>849</u>	<u>7.5</u>	<u>0.88</u>
<u>06/2012</u>	<u>947</u>	<u>02/2021</u>	<u>852</u>	<u>899.5</u>	<u>47.5</u>	<u>5.28</u>
<u>02/2012</u>	<u>896</u>	<u>02/2021</u>	<u>956</u>	<u>930.5</u>	<u>34.5</u>	<u>3.71</u>

Disclaimer/Publisher's Note: The statements, opinions and data contained in all publications are solely those of the individual author(s) and contributor(s) and not of MDPI and/or the editor(s). MDPI and/or the editor(s) disclaim responsibility for any injury to people or property resulting from any ideas, methods, instructions or products referred to in the content.